



Review

Bilateral cochlear implants in children: Effects of auditory experience and deprivation on auditory perception

Ruth Y. Litovsky^{a,*}, Karen Gordon^b^a University of Wisconsin-Madison, 1500 Highland Ave, Madison, WI, 53705, United States^b University of Toronto, Canada

ARTICLE INFO

Article history:

Received 7 October 2015

Received in revised form

7 January 2016

Accepted 11 January 2016

Available online 30 January 2016

ABSTRACT

Spatial hearing skills are essential for children as they grow, learn and play. These skills provide critical cues for determining the locations of sources in the environment, and enable segregation of important sounds, such as speech, from background maskers or interferers. Spatial hearing depends on availability of monaural cues and binaural cues. The latter result from integration of inputs arriving at the two ears from sounds that vary in location. The binaural system has exquisite mechanisms for capturing differences between the ears in both time of arrival and intensity. The major cues that are thus referred to as being vital for binaural hearing are: interaural differences in time (ITDs) and interaural differences in levels (ILDs). In children with normal hearing (NH), spatial hearing abilities are fairly well developed by age 4–5 years. In contrast, most children who are deaf and hear through cochlear implants (CIs) do not have an opportunity to experience normal, binaural acoustic hearing early in life. These children may function by having to utilize auditory cues that are degraded with regard to numerous stimulus features. In recent years there has been a notable increase in the number of children receiving bilateral CIs, and evidence suggests that while having two CIs helps them function better than when listening through a single CI, these children generally perform worse than their NH peers. This paper reviews some of the recent work on bilaterally implanted children. The focus is on measures of spatial hearing, including sound localization, release from masking for speech understanding in noise and binaural sensitivity using research processors. Data from behavioral and electrophysiological studies are included, with a focus on the recent work of the authors and their collaborators. The effects of auditory plasticity and deprivation on the emergence of binaural and spatial hearing are discussed along with evidence for reorganized processing from both behavioral and electrophysiological studies. The consequences of both unilateral and bilateral auditory deprivation during development suggest that the relevant set of issues is highly complex with regard to successes and the limitations experienced by children receiving bilateral cochlear implants.

This article is part of a Special Issue entitled <Annual Reviews 2016>.

© 2016 Elsevier B.V. All rights reserved.

Contents

1. Introduction	77
2. Spatial hearing in bilaterally implanted children	77
3. Direct stimulation of binaural hearing	80
4. Considerations regarding auditory plasticity	83
5. Summary	85
Acknowledgment	85
References	85

* Corresponding author.

E-mail address: litovsky@waisman.wisc.edu (R.Y. Litovsky).URL: <https://www.waisman.wisc.edu/bhl>

1. Introduction

As they grow, children are faced with increasing demands to function in complex listening environments, to integrate into mainstreamed educational and social situations, and to engage effectively and productively in a multi-sensory world. Spatial hearing skills are amongst the most important for maximizing these abilities, by providing critical cues for determining the locations of sources in the environment, and enabling segregation of important sources such as speech from background maskers or interferers. A child's ability to take advantage of these cues is likely to facilitate functioning in numerous everyday environments, enhancing incidental learning, and reducing fatigue and cognitive load.

It is well known that spatial hearing relies to some extent on monaural cues, but more importantly, good spatial hearing requires that the auditory system be able to integrate inputs arriving at the two ears from sound sources that vary in location in space. This “binaural integration” has exquisite mechanisms for capturing differences between the ears in time of arrival and intensity. One major cue thought to be vital for binaural hearing is the interaural difference in time (ITD), or difference between the ears in time of arrival of the sound. ITD cues in humans are equal to zero for sounds arriving from directly in front, and as sources are displaced laterally, ITDs grow more or less linearly until they reach a maximum value of approximately 700 μ s for an average adult size head. A second major cue is the interaural difference in level (ILD). This cue, which is equal to zero when stimuli arrive from directly in front, increases as stimuli are displaced laterally. ILDs result from ‘shadowing’ of the stimulus level at each ear, depend on the frequency of the sound being produced, such that the largest ILDs (–20 dB) occur at high frequencies. In addition, high frequency stimuli with amplitude modulations can contain ITD cues, and NH listeners show excellent sensitivity to such ITDs. CI speech processors are typically programmed with high-rate pulsatile stimulation; these rates are higher than the rates at which ITD sensitivity is observed. Thus, it is possible that envelope ITDs would be usable for spatial hearing abilities in CI users. However, the issue has not been studied in detail, and in fact data from NH listeners suggest that envelope ITDs have not been demonstrated to be effective for localization judgments (Macpherson and Middlebrooks, 2002).

In normally developing hearing systems, ITDs and ILDs are initially processed at the level of the auditory brainstem, and are continually refined and mediated by excitatory and inhibitory neuronal mechanisms that ultimately enable the brain to determine locations of sources in space. Excellent reviews on this topic are available for more detailed explanations and examples (Middlebrooks and Green, 1991; Blauert, 1997; Yin, 2002).

In children with normal hearing (NH), spatial hearing abilities are fairly well developed by age 4–5 years (Litovsky, 2011; Litovsky, 2015). However, it is likely that the ability to achieve good level of functioning regarding spatial hearing depends on the child's access to normal acoustic cues. By contrast, children who are deaf and hear through CIs do not have an opportunity to experience normal, binaural acoustic hearing early in life, and they function by having to utilize auditory cues that are degraded when it comes to numerous stimulus features (for review see Litovsky et al. (2012), Kan and Litovsky (2015)). When considering children who are bilaterally implanted, it is important to understand the extent to which maturation of spatial hearing depends on, and can vary with, exposure to bilateral stimulation. In addition, it is important to consider the potential role of experience on children's ability to utilize spatial information. For example, it is possible that binaural cues are weak or subtle, such that children require experience with them in order to learn to utilize them. However, if spatial cues are

not preserved with fidelity by the CI processors, then regardless of how much experience children have, they may never improve in their spatial hearing abilities. Additional studies are also discussed here, whereby the CI speech processors are bypassed, and research processors are used in order to provide binaural cues to specific pairs of electrodes in the two ears. Both behavioral and electrophysiological data are discussed here, to consider how place of stimulation as well level and timing of bilateral implant stimulation may affect spatial hearing in children. Finally, investigations of developmental plasticity after implantation are reviewed in an effort to explore both the successes and the limitations experienced by children receiving bilateral cochlear implants.

2. Spatial hearing in bilaterally implanted children

In the past 10–15 years there has been a progressive increase in the number of children receiving bilateral CIs, with growing evidence that, when listening with two CIs children generally perform better on spatial hearing tasks compared with unilateral listening modes. However, as is reviewed here, even with years of experience with bilateral CIs, most children do not perform as well as their NH peers.

A common measure of spatial hearing is sound localization. This ability has been studied using one of two approaches. First, localization acuity is a measure of how well a listener can discriminate between two source locations; often this measure is obtained for locations to the left vs. right, and the smallest angular difference between the two locations is a measure of acuity. A common metric is the minimum audible angle (MAA) (Litovsky, 1997; Hartmann and Raked, 1989). For recent reviews on this topic see also (Litovsky, 2011, 2015). A second measure is localization accuracy, which is informative regarding the ability of a listener to identify the location of a sound source from amongst an array of sources; accuracy can be measured for sound sources along the vertical or horizontal dimension but in bilateral CI users the focus has been on the horizontal plane, where binaural cues would be most effectively utilized (Zheng et al., 2015). Fig. 1 summarizes data from recent studies in which acuity and accuracy were measured in children with bilateral CIs. Of interest in these studies is the age of the children, task used and the amount of bilateral experience, which are summarized for panels 1A and 1B in Tables 1A and 1B, respectively.

As can be seen in Fig. 1, sound localization abilities of children with bilateral CIs have a number of important characteristics. First, the ability to discriminate sound sources, i.e., the MAA and other right-left discrimination measures, reveal a spatial hearing system that is fairly well established in a portion of the population. Some children as young as 2–3 years of age are able to discriminate left vs. right at small angles, with many children showing results that are within the range of what is seen in NH age-matched peers. However, there are also children who do not reach the performance level of age-matched peers and the reasons for this poor performance have not been fully identified (see further discussion below).

On the sound localization task, performance varies substantially across subjects, and it is possible that localization undergoes a more protracted developmental time course. Children who can typically perform the MAA task are not able to identify sound source locations well. While some of the children have error rates that are fairly low, none are as low as typically developing NH children. Fig. 2 shows examples of data from children between 5 and 14 years of age who received sequential bilateral CIs, and 5-year old children with NH. The best performing children with bilateral CIs have error rates within the range seen in the NH group. In the bilateral CI group, spatial hearing skills appear to be represented through different types of localization strategies that, with experience,




Fig. 1. A. Sound localization acuity as measured by MAA threshold for children using cochlear implants in bilateral and unilateral listening conditions as presented in three studies (NM = MAA was not measurable). Data come from Litovsky et al. (2006), Godar et al. (2010) and Grieco-Calub and Litovsky (2012; 96). B. Group mean RMS error in degrees presented in three studies comparing bilateral listening conditions to a unilateral listening condition and normal hearing control (* = No standard deviation reported for group mean RMS error). Data come from Grieco-Calub and Litovsky (2010), Van Deun et al. (2010) and Zheng et al. (2015).

Table 1A

Demographic information for studies shown in Fig. 1A.

	Litovsky et al., 2006	Godar and Litovsky, 2010	Grieco-Calub et al., 2012
Speaker locations	$\pm 70^\circ, \pm 60^\circ, \pm 50^\circ, \pm 40^\circ, \pm 30^\circ, \pm 20^\circ, \pm 10^\circ, 0^\circ$	$\pm 90^\circ, \pm 80^\circ, \pm 70^\circ, \pm 60^\circ, \pm 50^\circ, \pm 40^\circ, \pm 30^\circ, \pm 20^\circ, \pm 10^\circ, 0^\circ$	$\pm 70^\circ, \pm 60^\circ, \pm 50^\circ, \pm 40^\circ, \pm 30^\circ, \pm 20^\circ, \pm 10^\circ, 0^\circ$
Sample size	13 BiCI 6 UniCI	10 BiCI	27 BiCI 12 UniCI
Age (years)	3–16 yrs	5–10 yrs	2.2–3 yrs
BiCI Experience (months)	2–14 mos	Interval I: ~3 mos Interval II: ~12 mos	5.5–19.5 mos
Sequential/Simultaneous	All sequential	All sequential	16 sequential, 3 simultaneous
Manufacturer	All Cochlear Corp.	9 Cochlear Corp., 1 Advanced Bionics	All Cochlear Corp.

Table 1B

Demographic information for studies shown in Fig. 1B.

	Grieco-Calub and Litovsky, 2010	Van Duen et al., 2010	Zheng et al., 2015
Speaker locations	$\pm 70^\circ, \pm 60^\circ, \pm 50^\circ, \pm 40^\circ, \pm 30^\circ, \pm 20^\circ, \pm 10^\circ, 0^\circ$	$\pm 60^\circ, \pm 45^\circ, \pm 30^\circ, \pm 15^\circ, 0^\circ$	$\pm 70^\circ, \pm 60^\circ, \pm 50^\circ, \pm 40^\circ, \pm 30^\circ, \pm 20^\circ, \pm 10^\circ, 0^\circ$
Sample size	21 BiCI	30 BiCI	19 BiCI
Age (years)	5–15 yrs	4–15 yrs	4–9 yrs
BiCI Experience (months)	3–28 mos	12–44 mos	13–51 mos
Sequential/Simultaneous	All sequential	All sequential	16 sequential, 3 simultaneous
Manufacturer	17 Cochlear Corp., 3 Advanced Bionics, 1 Med-EL	All Cochlear Corp.	All Cochlear Corp.

undergo transitions from more rudimentary sound source categorization to more fine-grained ability to identify source locations. However, even after >4 years of bilateral experience children with bilateral CIs continue to show errors that are generally higher than the average age-matched NH peers (Zheng et al., 2015). Little work has been done to date on the impact of auditory training, but it is possible that localization abilities would be further improved by training in a spatial hearing task.

Overall the data also show that the bilaterally implanted children do not reach the same level of performance as their age-matched NH peers. In fact, even the bilateral CI users who have had >6 years of experience listening with their CIs did not perform as well as the children with NH. There are numerous factors that might account for the gap in performance between children with bilateral CIs and NH children: (1) When clinical processors are used

and sounds are presented to the children from loudspeakers in a sound field, there is no coordinated activation of the two speech processors. (2) CI processors do not deliver temporal fine structure cues that are important for spatial hearing using low frequency sounds. (3) It is possible that there are anatomical mis-matches between the depth of insertion of electrodes in the right and left ears. This might create a mis-match in the frequency representation of signals presented to electrodes along the arrays in the two ears that are anatomically parallel. This could minimize the ability of the binaural circuits to integrate inputs from the two ears and compute source locations. These factors are also discussed in relation to adults with bilateral CIs in a recent review (Kan and Litovsky, 2015). Effects of asymmetric hearing during development may also be relevant as discussed further below.

A second measure of spatial hearing is known as “spatial release




Fig. 2. Individual children's root mean square (RMS) error values are shown for children with normal hearing (NH) in panel A and children with bilateral CIs in panel B. NH children were 5 years old and CI users ranged in age from 5 to 14. For each group, data are sorted from lowest to highest errors along the x-axis. In both panels, data are arranged along the x-axis from the smallest to largest RMS errors (better to worse performance). In the bilateral group, the letters in the parentheses under each subject's data point indicate the subgroup that each child belonged to based on performance. In group A children had performance similar to that of NH children. In group B children were able to discriminate between the right and left but had poor ability to localize within each hemifield. Finally children in group C had poor ability to localize or discriminate between right and left. The dotted lines represent chance performance for either a 7-loudspeaker or a 15-loudspeaker setup, respectively (this is Fig. 2 from Grieco-Calub and Litovsky, 2010).




Fig. 3. Spatial configurations used in the studies on spatial release from masking. From left to right, panels show conditions with target speech in front, and interferers either in front, distributed asymmetrically around the head or distributed symmetrically around the head.

from masking" (SRM), which refers to the improvement in speech understanding in the presence of interfering sounds, also known as maskers. Testing is conducted when the target speech and maskers are spatially separated vs. when they are co-located. This improvement can either be measured by comparing changes in percent correct for speech understanding, or by measuring differences in speech awareness or reception thresholds for separated vs. co-located conditions. Fig. 3 shows a schematic of conditions used to test for the SRM effect. In the co-located condition, the target and maskers are presented from the front, and signals at the two ears have the time arrival intensity. In the asymmetrically separated condition, the target in front but the maskers are presented from one side, thus the ear farther from the masker has a reduced overall level relative to the ear that is near the masker. The target thus has a good signal-to-noise ratio (SNR) in the ear that is far from the masker (left ear in this rendition) compared with the ear near the masker. This 'better ear effect' enables listeners to rely on monaural head shadow cues for extracting information from the target speech. By contrast, in the symmetrically separated condition, having maskers on both sides of the head renders neither ear in a position of having a good SNR. In this case, listeners do not have monaural head shadow cues, and must rely on binaural processing for source segregation.

Fig. 4 shows recently published data (Misurelli and Litovsky, 2015) in which these conditions were studied and results are

compared for the bilateral CI users and NH children as well as NH adults.

Results suggest that NH children had SRM in both conditions; however, the effect was larger in the asymmetrical condition. Thus, having access to normal acoustic cues renders children able to use binaural cues for source segregation when monaural head shadow cues are minimal or absent. In addition, the monaural head shadow cues do provide an added benefit for NH children. In the CI groups SRM was small or even negative in the symmetrical conditions, and measurable, but smaller than that of NH children in the asymmetrical conditions. This findings suggested that children who use bilateral CI rely on the ear with a better SNR, however, binaural mechanisms involved in SRM are weak or absent in these children. Results from a study by Van Deun and colleagues (Van Deun et al., 2010) are consistent with this observation; bilaterally implanted children demonstrated benefits due to head shadow effects around 4–6 dB, which was comparable to effects observed in NH children. However, effects due to binaural processing were not observed.

A second effect studied was related to target-masker similarity. It has been suggested that spatial cues become especially useful for source segregation when sources are highly similar or confusable (Durlach et al., 2003). Similarity and confusability between target and maskers can occur, for instance, when the sources consist of same-sex vs. different-sex talkers (e.g., Brungart et al., 2001). SRM was larger in conditions with same sex talkers (i.e., in the top panel where both target and masker were male), a finding that is consistent with observations reported in NH adults (Jones and Litovsky, 2011).

The children with CIs had the same 'hearing age' as the NH children, suggesting that auditory experience is not sufficient for explaining why the gap in performance exists between bilaterally implanted children and NH children. Further evidence comes from data in NH children who show large SRM values by 3–5 years of age (Garadat and Litovsky, 2007; Litovsky, 2005). Recently, Misurelli and Litovsky (2015) found that when SRM is measured at annual intervals, the effects are not stable across the population of children with bilateral CIs: in some children SRM increased with additional listening experience, in other children the reverse was true, and in others there were no notable effects. Because test-retest data from a previous study showed that performance was stable (Misurelli




Fig. 4. SRM values are plotted for normal hearing (NH) adults, and for groups of children with different ages, with either NH or with bilateral CIs. A and B refer to younger or older, respectively. The chronological ages of the NH groups were on average 4 years 9 months (NH-A) and 5 years 2 months (NH-B). Ages were 7 years 5 months and 8 years 3 months for the BiCI-A and BiCI-B groups, respectively. These data are from Misurelli and Litovsky (2015).

and Litovsky, 2012), the changes denoted here were discussed in the context of other factors. For example, the authors speculated about the possible detrimental effects caused by signal processing in the microphones, about lack of bilateral mapping and hence asymmetries between frequency-allocation of speech signals in the two ears. They also discussed the possibility that children who use FM systems may become accustomed to listening to speech that has excellent SNR, and are therefore less adept at extracting information from target speech in the presence of maskers when the FM system is not available.

Asymmetric auditory development impairs speech unmasking. Children with simultaneous implantation of bilateral CIs benefited from moving noise away from speech by 90° by 7.2 dB with no effect of the side of noise in measures of speech unmasking whereas a group of children who were sequentially implanted performed significantly better when noise was on the side of the second ear (4.8 dB) rather than the first ear (3.0 dB) (Chadha et al., 2011). The asymmetric findings in the sequentially implanted

group were consistent with earlier reports (Mok et al., 2007), showing weaker effects (<1 dB) on speech detection thresholds when noise was moved to the side of the first CI in bilaterally implanted children and to the CI in children using a hearing aid in the other ear (bimodal users) than when noise was moved toward the poorer hearing/second implanted ear (~4 dB). Speech reception threshold data from Sparreboom and colleagues (2011) show a similarly small (1.8 dB) advantage of moving noise from the co-incident position with speech in front to the side of the first CI. Effects of moving noise to the opposite side of the head were not measured and thus asymmetry was not assessed in that study. More recently, Killan et al. (2015) have shown that SRM measured with speech reception thresholds improves with ongoing bilateral implant use between 2 and 4 years following sequential cochlear implantation. This is consistent with the finding that SRM decreases with age in bilateral implant users (Chadha et al., 2011). Poorer benefits when noise was moved to the first than second implanted ear were found at 24 months of bilateral implant use (3.3 dB poorer) (Killan et al., 2015). The asymmetry decreased but persisted at a follow-up of 48 months of bilateral use (1.8 dB asymmetry in SMR). In sum, the data indicate that children using bilateral cochlear implants obtain spatial benefits for listening to speech in noise but the advantages can be limited by asymmetric development and other mismatches between devices.

3. Direct stimulation of binaural hearing

Studies discussed thus far present stimuli from loudspeakers and allow the speech processors in the two ears to respond to the stimuli as they would in clinical situations. Because the two CIs are not coordinated, as described above, binaural cues are degraded, discarded or possibly occur in a spurious fashion. Studies using research processors provide an opportunity to directly stimulate select pairs of electrodes in the two ears, and the researcher can control many important parameters. Both behavioral and electrophysiological data in bilateral CI users reveal that the ability of CI users to integrate stimuli from the two depends on the place of stimulation along the cochlear implant electrode arrays, the level of the stimuli at each electrode and the relative timing of the signals. Most of the work on this topic to date has been conducted in groups of adult CI users [reviewed in Kan and Litovsky (2015), Laback et al. (2015)]. The current review focuses on what has been learned to date in children, and references the adult literature for comparison. Overall, findings with adults suggest that, despite tight control over binaural cues with research processors, fine-structure ITD sensitivity is poorer than that seen in NH adults. More so, ITD sensitivity in CI users is best at very low pulse rates, and very poor at higher pulse rates, i.e., rates that are typically used in clinical speech processors.

Measures of binaural sensitivity in children with bilateral CIs reveal important differences from their NH peers. Litovsky and colleagues use identical paradigms to those established to date with the adult patients, except that the interactive computer tools are designed to engage the children in the tasks, reinforce them throughout the experimental procedures. In these studies, pulsatile stimulation is presented with varying rates and great effort is taken to find electrodes in the two ears that are matched by pitch and perceived loudness (e.g., Kan et al., 2013; Litovsky et al., 2010).

One common perceptual measure of binaural sensitivity is discriminability for sounds that vary in either ITD or ILD. In a recent study, Litovsky and colleagues (Ehlers et al., 2015) used two types of stimuli to measure binaural sensitivity in 8–10 year old NH children. One stimulus was a transposed tone, which has high-rate carriers with amplitude modulations. The other was a CI simulation using a Gaussian Envelope Tone (GET) vocoder. In both cases,

ITD cues are available in the envelope but not in the temporal fine structure. On average, NH children were able to discriminate ITDs and ILDs as well as adults, but there was more variability in the children than adults. ITD Thresholds in children ranged from 34 to 389 μ s and 15.6–362 μ s for the transposed and GET stimuli, respectively. In adults, thresholds ranged from 38 to 124 μ s and 27–356 μ s for the transposed and GET stimuli, respectively. ILD thresholds in children were 0.5–1.8 and 0.7–6 dB for transposed and GET stimuli, respectively. Finally, in adults ILD thresholds 0.5–1.8 and 0.5–6.7 dB, for the transposed and GET stimuli, respectively. Overall, those findings suggest that binaural sensitivity to ITDs in the envelopes is achievable by children ages 8–10 with NH, and that auditory development is unlikely to be the reason for poor performance in bilaterally implanted children.

In that same study, children's ability to perceptually locate sounds "in the head" (lateralization) was also measured, and results showed a markedly different level of performance in children vs. adults, suggesting that, although binaural acuity for these stimuli is well developed in early childhood, there is poor spatial mapping of those same cues. This finding has similarities to the results described above from the MAA and localization studies in free field, in that children with bilateral CIs are generally capable of discriminating source locations to the right vs. left, but they have a poorly established map of auditory space.

In general, it appears important to provide stimulation at places of stimulation that are matched because the bilateral pathways are tonotopically organized (Grothe et al., 2010). Yet, the most commonly used cochlear implant stimulation mode is monopolar where the current can spread over a range of several mm, stimulating large regions of the auditory nerve in each ear. Thus, it is not clear to what extent exact matching of stimulation across the ears by place and/or pitch perception is needed. Gordon and colleagues (2012a) showed that children who are deaf retain tonotopic organization of bilateral pathways. In that study, the authors used an electrophysiological measure of the auditory brainstem called the binaural interaction component (BIC) or binaural difference (BD) (Gordon et al., 2012a). This measure, also used to match electrodes in bilaterally implanted cats (Smith and Delgutte, 2007) and adult users (He et al., 2010, 2012), produces a difference between the sum of brainstem responses to each implant separately and the response to the same bilaterally presented input. Importantly, the BD was no longer present when stimulation with large mismatches in place were delivered to children with full insertions of both devices (Gordon et al., 2012a). Bilateral stimulation with smaller mismatches in place had little effect on the BD at comfortably loud current levels (Gordon et al., 2012a) although slightly greater effects were noted at lower levels in adult cohorts (He et al., 2010, 2012). Similar to the electrophysiological results, there appears to be some tolerance in behavioral data for slight mismatches in place of stimulation between implants; binaural sensitivity is still observed for cases in which mis-matches of up to ~2–3 mm occur (Goupell et al., 2013; Kan et al., 2015).

The electrophysiological BD study was also helpful in assessing binaural coding of ILDs in children. Significant changes in BD amplitude were found with increases in ILD, suggesting a change in bilateral brainstem activity (Gordon et al., 2012a). Moreover, children increasingly lateralized these sounds to one side of their head as BD amplitude increased. In later work, Gordon and colleagues found that perception of ILDs was established at very early stages of bilateral implant use even for children who were older and heard through one implant for most of their lives (Gordon et al., 2014). This was consistent with findings that most children are able to discriminate stimuli containing ILD cues when tested after longer periods of bilateral implant use (Ehlers et al., 2015; Salloum et al., 2010). More recently, ILD sensitivity in the envelope of short

pulse trains in CU was shown to be slightly reduced in children with long bilateral implant use (simultaneously and sequentially implanted) compared to normal hearing peers listening to clicks (dB re: 20 Pa) (Gordon et al., 2014) (Fig. 5A).

By contrast, studies using electrically pulsed signals reveal that children's sensitivity to ITDs is weak relative to their ILD sensitivity. Indeed, Gordon and colleagues' initial impression was that ITD sensitivity was absent in most children with bilateral implants (Salloum et al., 2010) with similar findings from Ehlers et al. (2015). Weak detection of ITDs was recently reported by Gordon et al. to emerge in children with over 4 years of bilateral implant experience (Gordon et al., 2014). The sensitivity to these cues was significantly poorer than normal as shown in Fig. 5B. Both responses to ILD and ITD were fit with logit regression curves and the mid-point (0.5) calculated for both cues. The linear relationships plotted for each of 3 groups of children indicated that when ITDs were presented with levels weighted to one side, ITD shifts to the opposite side were required to achieve balanced percepts. This time-intensity trading was very specific for children with normal hearing listening to acoustic clicks (27 μ s shift in ITD for every 1 dB of ILD). Although the direction was similar for bilaterally implanted children listening to single electrical pulses, the shifts in ITD were relatively large (97–153 μ s for every 1 CU shift in ILD). This means that ITD sensitivity is much weaker than normal in children with bilateral implants. These data reflect the typical exposure to binaural cues by children using bilateral CIs. Whilst they have access to ILDs through their bilateral devices, they continue to lack consistent exposure to ITDs in daily listening, thereby deteriorating their ability to detect these cues when provided under controlled experimental conditions through direct stimulation. Yet, the emergence of weak ITD detection with long bilateral implant experience suggests that children could be developing unique strategies of listening to bilateral electrical input under both direct experimental conditions as well as through their devices, consistent with Zheng et al. (2015), as discussed above).

An example of unique binaural listening of direct electrical stimulation comes from a study from Gordon and colleagues (Steel et al., 2015), in which bilaterally implanted children were reported to inconsistently hear a fused bilateral image in experimental conditions. As shown in Fig. 6A, children frequently indicated that they heard 1 rather than 2 sounds when listening to simultaneous pulse trains from paired apical implant electrodes containing ILDs. As discussed above, these inputs were likely lateralized to one side as shown in similar cohorts of children (Gordon et al., 2014; Salloum et al., 2010). However, when bilateral input was presented with ITDs at balanced levels (ie. no ILDs), children with bilateral implants responded at chance levels when asked if they heard 1 or 2 sounds (Steel et al., 2015) (Fig. 6B). This means that, without ILD cues, direct bilateral cochlear implant input is difficult for children to fuse into one image. The lack of ITD lateralization at early stages of bilateral device use under similar conditions could thus be explained in part by an inability to fuse bilateral input. If this poor binaural fusion continues despite long term bilateral implant experience, the later emergence of ITD detection reported above (Gordon et al., 2014) might reflect an abnormal strategy used by these children to detect long ITDs. Such strategies may not be useful for shorter ITDs more typically needed for spatial listening. In the same study, reaction time and pupilometry were used to gauge listening effort involved during the binaural fusion task across groups (Steel et al., 2015). Clear increases in both of these measures occurred with decreases in perception of a single fused binaural image. This suggests that mismatches in hearing between the two ears make listening more effortful for both children with normal hearing as well as for children with bilateral implants. It may therefore be important to correct for inaccurate binaural cues




Fig. 5. A. Experienced bilateral CI users and normal hearing peers perceived changes in ILDs as these cues moved from left to right weighted. Interaural level differences (ILDs) in CI users represent differences in CU and dB re: 20 Pa in the normal hearing group (0.08 dB re: 100 mPa per acoustic dB change in ILD.) (Mean data \pm 1SE from Fig. 3 from Gordon et al., 2014). B. Time-intensity trading revealed by plotting predicted balance (ie. 0.5 responses calculated from logit regressions of responses) to ILD and ITD cues. Positive values are right weighted ILDs and left leading ITDs. ILDs weighted to one side shifted balanced perception toward opposite leading ITDs in all groups. Time-intensity trading shifts were much larger for the two groups of bilaterally implanted children than the normal hearing peer group (Fig. 6 from Gordon et al., 2014).

by improving programming of bilateral implants and better coordinating the speech processors as previously suggested (Litovsky et al., 2012; Kan et al., 2013; Gordon et al., 2012b).

A related finding with implications for binaural sensitivity comes from measures of binaural masking level differences (BMLDs), which is akin to SRM, in that the task involves segregation of target from noise through use of binaural cues. Van Deun et al. (201) presented children with encoded temporal information at one pair or multiple pairs of electrodes, and compared performance with either diotic (target and noise having the same binaural cues), or dichotic (target and noise presented with different ITD values) conditions. In the BMLD paradigm, binaural benefits of dichotic condition are thought to arise from a de-correlation of the signals between the two ear, which can produce dynamic ITDs and ILDs that are potentially available for extracting information about the target in the presence of the noise. More specifically, interaural de-correlation might be providing listeners with cues regarding the spectral structure of

the target. In this study, children demonstrated BMLDs of a few dB, suggesting they were sensitive to interaural de-correlation of temporal envelope in dichotic listening conditions. As reported above, because many children with bilateral CIs are weakly sensitive to ITDs if at all, this finding indicated that the unmasking conditions revealed sensitivity to dynamic interaural level differences (ILDs), or interaural de-correlation of temporal envelope cues.

An interesting issue regarding studies with single pairs of electrodes is that it is difficult to interpret how they will compare to speech processing with multiple channels of stimulation. In another recent study Todd et al. (2015), binaural unmasking was measured with multiple and single pairs of electrodes. Children with bilateral CIs performed similarly in the multi-electrode conditions and in the single-electrode conditions, and their performance pattern was similar to what was found with NH children who listened to simulations of CIs. Interestingly the results were not predicted from whether those children had good vs. poor ITD




Fig. 6. A: Children with bilateral implants ($n = 25$) consistently perceived one image when presented with unilateral (Level in CI-1 = 0, CI-2 at comfortable loud level) and bilateral input (increasing CI-1 level, CI-2 at comfortable loud level) albeit less frequently than did NH peers ($n = 24$) ($p < 0.0001$). 6B) Children with bilateral implants respond at chance to bilateral balanced stimuli with balanced input and significantly less often than the normal group ($p < 0.0001$). When ITDs are extremely large (24 μ s), both groups indicate a lack of binaural fusion, demonstrating understanding of the task (Mean data \pm 1 SE from Fig. 5, Steel et al., 2015).

sensitivity, suggesting that the auditory cues used for binaural unmasking and binaural spatial hearing may not be the same. Overall, one of the main advantages gained from binaural unmasking is the improved ability of bilaterally implanted children to hear speech in noise at lower (worse) SNRs.

4. Considerations regarding auditory plasticity

As reviewed above, children with early onset deafness use their bilateral cochlear implants differently than adults. Although neither of these bilaterally implanted groups receives access to accurate binaural cues (as reviewed above), the increasingly abnormal findings in children suggests a unique vulnerability of the auditory system to the loss of normal hearing from both ears in early development. This finding might appear contrary to the notion of increased plasticity in children than adults. Yet, as will be discussed here, large developmental changes observed along the auditory pathways with implant use in children do not always occur in expected directions. The differences from normal reflect the atypical input provided by cochlear implants to children with hearing loss.

For many children, the loss of normal bilateral input begins during the period of deafness and continues while they use a unilateral cochlear implant. Development is driven preferentially from the unilaterally stimulated ear (Gordon et al., 2015), compromising responses from the opposite pathways and potentially disrupting binaural hearing. Efforts to avoid these changes include simultaneous bilateral implantation (Summerfield et al., 2002; Ramsden et al., 2009; Henkin et al., 2014; Lopez-Torrijo et al., 2015), and the use of a hearing aid in the unimplanted ear (bimodal device use) (Mok et al., 2010; Nitttrouer and Chapman, 2009; Ching et al., 2009; Polonenko et al., 2015).

Children's poor ability to fuse direct electrical stimulation from pairs of bilateral CI electrodes may be particularly revealing. Perception of these inputs as one sound deteriorated with increasing asymmetry of brainstem function as measured by differences in brainstem response latencies (Steel et al., 2015). These asymmetries arise from sequential stimulation of one ear before the other (Gordon et al., 2012a; Gordon et al., 2007a; Gordon et al.,

2007b; Gordon et al., 2008; Sparreboom et al., 2010). Of importance, the eIII-eV latency (neural conduction time from cochlear nucleus to midbrain) (Gordon et al., 2006) remained arrested in children with early onset deafness across a wide range of ages with little stimulation from either side prior to implantation (Gordon et al., 2006, 2003; Gordon et al., 2010). Rapid decreases in the eIII-eV were then measured over the first 6 months of unilateral cochlear implant use (Gordon et al., 2006, 2003, 2007c; Thai-Van et al., 2007). Thus, it appears that the persistent immaturity of the bilateral brainstem pathway allows plasticity to be retained during bilateral deafness. Unilateral stimulation drives the brainstem toward maturation, possibly closing a sensitive period of developmental plasticity and leaving brainstem pathways from the opposite ear relatively immature (as measured by delayed latencies from the second implanted side (Gordon et al., 2012a; Gordon et al., 2007a; Gordon et al., 2007b; Gordon et al., 2008; Sparreboom et al., 2010)). Unilaterally stimulated brainstem neurons will also be deprived of bilateral input, which is primarily inhibitory from the ipsilateral side (Grothe et al., 2010). Fortunately, these connections are retained, as shown by clear binaural differences (BDs) in brainstem responses in both children (Gordon et al., 2012a) and adults (He et al., 2010) with bilateral CIs. This means that binaural processing in the brainstem may be altered by unilateral implant use but not obliterated. Asymmetric timing in the brainstem is in agreement with functional findings; although children's sensitivity to small ILDs (Litovsky et al., 2012; Gordon et al., 2014) emerges very quickly after bilateral implantation (Gordon et al., 2014), detection of timing cues are relatively poor (Litovsky et al., 2012; Litovsky et al., 2010; Gordon et al., 2014; Salloum et al., 2010), take many years to emerge (Gordon et al., 2014), and appear to be weighted to the first implanted ear (Gordon et al., 2014) as discussed above. Some of these listening challenges may contribute to or even result from poor binaural fusion of direct electrical stimulation (Steel et al., 2015) as also discussed above.

The degree of asymmetry in brainstem function is generally in line with a number of behavioral findings demonstrating that children listen better with their first than second implanted ears (Illg et al., 2013; Fitzgerald et al., 2013; Gordon and Papsin, 2009). Yet, these associations could also reflect changes mediated in more




Fig. 7. A–D, Schematic representations of the strength of pathways from each ear (right ear in red, left ear in blue) to the contralateral and ipsilateral auditory cortices are shown. E, Mean (SE) dipoles measured in the left (blue) and right (red) auditory cortices (Gordon et al., 2013). Responses to pure tones in each ear are stronger in the contralateral than ipsilateral cortex in children with normal hearing. Bilateral pathways are essentially symmetric in children receiving bilateral input in early development (normal-hearing, simultaneous bilateral implants, short delay between implants). Unilateral right implant use strengthens pathways to both contralateral and ipsilateral auditory cortices, increasing dipoles in the left auditory cortex and reversing aural preference to the first implanted right ear. CI-1, first implant; CI-2, second implant (Fig. 4 from Gordon et al., 2015).

central areas of the auditory system (Gordon et al., 2013). Cortical activity measured in dipole moment (nAm) by electroencephalography and located in left and right auditory hemispheres revealed reorganization with unilateral implant use as summarized in Fig. 7 from 4 groups: 1) normal hearing children, 2) children using a right unilateral cochlear implant, and two groups of children receiving bilateral implants 3) simultaneously implanted or 4) sequentially implanted after use of a right sided cochlear implant. The sequentially implanted group was further divided by analyses of cortical activity by duration of unilateral implant use. Responses from localized sources in auditory cortices became significantly different from the simultaneous group by ~1.5 years, thus providing a cut point for the sequential group into “Long Delay” and “Short Delay” groups. Activity in both auditory cortices was strengthened from the right implant in the unilateral and Long Delay groups supporting the concept of an “aural preference” for the first hearing ear when bilateral implant is delayed for >1.5 years. In the normal hearing group, significantly contralateral responses for each ear were found and each cortex showed significant contralateral aural preference (7A). These findings were consistent with data in animal models (Kral et al., 2013), and likely reflect cortical immaturity in this cohort (Jiwani et al., 2016). An increased response in the left cortex was measured in the group with unilateral right implants (7B). Simultaneous bilateral implantation protected the normal pattern of responses (7C) while children in the Long Delay showed both an increased left auditory cortex response as well as an

abnormal aural preference for the ipsilateral (first hearing right ear) in the right auditory cortex (7D). Mean (SE) dipoles in each cortex stimulated by each ear for all groups are shown (7E).

It is difficult to distinguish cortical from brainstem contributions to behavioral findings. Reorganization of cortical activity may be promoted from a lack of contralateral inhibition to bilateral brainstem nuclei during the period of unilateral listening but it is also possible that the cortex reorganizes in response to altered binaural cues ascending through a brainstem which became preferentially driven from one side. Moreover, the role of descending projections from the cortex to midbrain and brainstem is unknown. Neuroanatomical reorganization of auditory cortex has also been found in children with asymmetric or unilateral deafness (Propst et al., 2010; Schmithorst et al., 2005; Schmithorst et al., 2014; Tibbetts et al., 2011; Vasama and Makela, 1997), confirming the influence of unilateral deprivation in development and further supporting the importance of early intervention (Gordon et al., 2015; Lieu et al., 2010; Bess and Tharpe, 1984).

Asymmetries in auditory activity and function persist with several years of bilateral implant use in children, suggesting more limited plasticity of the newly implanted ear than that of the first implanted ear. Whether development in the second implanted ear can ultimately “catch up” with the other remains unanswered. It may seem reasonable to define an age at which bilateral implantation should not proceed. However, while benefits from bilateral implantation in children can decline with age (Van Deun et al.,

2010; Fitzgerald et al., 2013; Strom-Roum et al., 2012; Galvin and Hughes, 2012; Scherf et al., 2009a; Scherf et al., 2009b; Wolfe et al., 2007; Peters et al., 2007; Graham et al., 2009), there are many associated time-based issues to address. Consider for example, the suggestion by Graham and colleagues that adolescents should not be provided with a second cochlear implant (Graham et al., 2009). The adolescent group may be amongst the poorest candidates based on abnormal cortical maturation, poor speech perception in the newly implanted ear compared to the first, inconsistent use of the second device, and inability to hear binaural timing cues (Gordon et al., 2014; Graham et al., 2009; Jiwani et al., 2016). Yet, several time-based factors might explain these findings including long durations of deafness experienced by the unimplanted ear which might decouple it from the secondary auditory cortices (Kral, 2013); a lack of bilateral hearing/duration of unilateral hearing which is well beyond the optimal “Short Delay” of 1.5 years (Gordon et al., 2015, 2013); a relatively mature stage of cortical development at which the auditory stimulation is being introduced (Jiwani et al., 2013; Kral and Sharma, 2012; Ponton and Eggermont, 2001) and; finally, the stage of teenage life at implantation which comes with unique social and educational pressures (Borman et al. 1998; Blakemore and Mills, 2014). Likely, all of these issues are important contributory factors which means that age alone cannot provide a meaningful explanation for the poorer outcomes of bilateral implantation demonstrated in some adolescent cohorts.

Access to hearing in early life is clearly important for development of the auditory cortex (Kral, 2013) and spoken language (Geers and Nicholas, 2013) and, as reviewed above, is also essential for binaural hearing. With that in mind, the type of bilateral input needed in early development is important to address. Gordon and colleagues have been exploring whether acoustic stimulation in an ear with some residual hearing can be integrated with a cochlear implant in the other ear to preserve the bilateral auditory system and promote binaural hearing (Mok et al., 2010; Nitttrouer and Chapman, 2009; Ching et al., 2009; Polonenko et al., 2015). Present studies are also investigating effects of implantation in children with unilateral deafness who in the past were not cochlear implant candidates and were thus unable to access binaural cues (Vincent et al., 2015). Again, time-based factors need to be considered as many of these children experience progressive deterioration of hearing (Polonenko et al., 2015). This means that they experience different durations of deprivation either unilaterally or bilaterally depending on their history of hearing loss identification and treatment including time to implantation. All are coupled with advancing age and stage of auditory development. Early findings by Arndt and colleagues are consistent with the importance of bilateral input in childhood as shown by poorer outcomes in both adults and children with long durations of early onset unilateral deafness (Arndt et al., 2015).

Efforts to preserve bilateral auditory function in both children and adults rest on the notion that auditory processing of bilateral input through bilateral implants or bimodal devices, while not normal, might provide benefits relative to adapting to unilateral hearing. Abnormalities in secondary cortical areas (Schmithorst et al., 2014; Tibbetts et al., 2011; Jiwani et al., 2016) for auditory processing coupled with findings of increased effort, poorer educational and speech language development (Lieu et al., 2010; Bovo et al., 1988; Lieu, 2013; Lieu et al., 2012) in children with unilateral hearing, and benefits of both bilateral and bimodal use over unilateral implant use in children (as discussed above, Fig. 1, and (Giannantonio et al., 2015; Ching et al., 2014)) suggest that this has been the right move. Nonetheless, future efforts to improve binaural integration and access to binaural timing cues will make better use of the bilateral pathways we are aiming to preserve by stimulating bilaterally during early development.

5. Summary

Spatial hearing skills are fundamental for the ability of children to function in complex auditory environments. The emergence of these skills during development is clearly hampered in children with bilateral CIs, and the reasons for these limitations are complex. The clinical approach is to attempt to maximize binaural sensitivity in order to improve spatial hearing, but the ability to present true binaural cues through processors is limited by today's technology. If that were possible, however, early exposure to appropriate cues would have the potential to help these children function more similarly to their NH peers. In addition, implantation varies with regard to whether children receive their CIs simultaneously or sequentially. The ability of the brain to reorganize is greatly affected by the early exposure that children have to appropriate cues, and it is clear that there are consequences of both unilateral and bilateral auditory deprivation during development.

Acknowledgment

The authors acknowledge funding from the National Institutes of Health (NIH-NIDCD GRANTS 5R01 DC008365 and 5R01 DC03083 to RYL) and Canadian Institutes of Health Research (MOP-130420, MOP-97924, MOP-89804) and the Hearing Foundation of Canada to KG. The authors are grateful to Rachael Joicevicz for assistance with preparation of figures and tables.

References

- Arndt, S., et al., 2015. Cochlear implantation in children with single-sided deafness: does aetiology and duration of deafness matter? *Audiol. Neurotol.* 20 (Suppl. 1), 21–30.
- Bess, F.H., Tharpe, A.M., 1984. Unilateral hearing impairment in children. *Pediatrics* 74 (2), 206–216.
- Blakemore, S.J., Mills, K.L., 2014. Is adolescence a sensitive period for sociocultural processing? *Annu. Rev. Psychol.* 65, 187–207.
- Blauert, J., 1997. *Spatial Hearing*. MIT, Cambridge, MA, p. 494.
- Borman, K.M., Schneider, B.L., National Society for the Study of Education, 1998. *The Adolescent Years : Social Influences and Educational Challenges. Ninety-seventh Yearbook of the National Society for the Study of Education. NSSE : Distributed by the University of Chicago Press, Chicago*, p. 240 xv.
- Bovo, R., et al., 1988. Auditory and academic performance of children with unilateral hearing loss. *Scand. Audiol. Suppl.* 30, 71–74.
- Brungart, D.S., et al., 2001. Informational and energetic masking effects in the perception of multiple simultaneous talkers. *J. Acoust. Soc. Am.* 110 (5 Pt 1), 2527–2538.
- Chadha, N.K., et al., 2011. Speech detection in noise and spatial unmasking in children with simultaneous versus sequential bilateral cochlear implants. *Otol. Neurotol.* 32 (7), 1057–1064.
- Ching, T.Y., et al., 2009. Bimodal fitting or bilateral implantation? *Cochlear Implants Int.* 10 (Suppl. 1), 23–27.
- Ching, T.Y., et al., 2014. Language and speech perception of young children with bimodal fitting or bilateral cochlear implants. *Cochlear Implants Int.* 15 (Suppl. 1), S43–S46.
- Durlach, N.I., et al., 2003. Note on informational masking. *J. Acoust. Soc. Am.* 113 (6), 2984–2987.
- Ehlers, E., Zheng, Y., Goupell, M., Godar, S.P., Litovsky, R.Y., 2015. Measuring binaural sensitivity in children who use bilateral cochlear implants. *J. Acoust. Soc. Am.* (submitted for publication).
- Fitzgerald, M.B., et al., 2013. Factors influencing consistent device use in pediatric recipients of bilateral cochlear implants. *Cochlear Implants Int.* 14 (5), 257–265.
- Galvin, K.L., Hughes, K.C., 2012. Adapting to bilateral cochlear implants: early post-operative device use by children receiving sequential or simultaneous implants at or before 3.5 years. *Cochlear Implants Int.* 13 (2), 105–112.
- Garadat, S.N., Litovsky, R.Y., 2007. Speech intelligibility in free field: spatial unmasking in preschool children. *J. Acoust. Soc. Am.* 121 (2), 1047–1055.
- Geers, A.E., Nicholas, J.G., 2013. Enduring advantages of early cochlear implantation for spoken language development. *J. Speech Lang. Hear. Res.* 56 (2), 643–655.
- Grieco-Calub, T., Litovsky, R.Y., 2012. Spatial acuity in two-to-three-year-old children with normal acoustic hearing, unilateral cochlear implants and bilateral cochlear implant. *Ear Hear* 33 (5), 561–572.
- Giannantonio, S., et al., 2015. Experience changes how emotion in music is judged: evidence from children listening with bilateral cochlear implants, bimodal devices, and normal hearing. *PLoS One* 10 (8) e0136685.
- Godar, S.P., Litovsky, R.Y., 2010. Experience with bilateral cochlear implants improves sound localization acuity in children. *Otol. Neurotol.* 31 (8), 1287–1292.

- Gordon, K.A., et al., 2008. Abnormal timing delays in auditory brainstem responses evoked by bilateral cochlear implant use in children. *Otol. Neurotol.* 29 (2), 193–198.
- Gordon, K.A., et al., 2010. Auditory development in the absence of hearing in infancy. *Neuroreport* 21 (3), 163–167.
- Gordon, K.A., et al., 2012a. Binaural interactions develop in the auditory brainstem of children who are deaf: effects of place and level of bilateral electrical stimulation. *J. Neurosci.* 32 (12), 4212–4223.
- Gordon, K.A., et al., 2012b. Toward a method for programming balanced bilateral cochlear implant stimulation levels in children. *Cochlear Implants Int.* 13 (4), 220–227.
- Gordon, K.A., et al., 2014. Perception of binaural cues develops in children who are deaf through bilateral cochlear implantation. *PLoS One* 9 (12), e114841.
- Gordon, K.A., Papsin, B.C., 2009. Benefits of short interimplant delays in children receiving bilateral cochlear implants. *Otol. Neurotol.* 30 (3), 319–331.
- Gordon, K.A., Papsin, B.C., Harrison, R.V., 2003. Activity-dependent developmental plasticity of the auditory brain stem in children who use cochlear implants. *Ear Hear.* 24 (6), 485–500.
- Gordon, K.A., Papsin, B.C., Harrison, R.V., 2006. An evoked potential study of the developmental time course of the auditory nerve and brainstem in children using cochlear implants. *Audiol. Neurotol.* 11 (1), 7–23.
- Gordon, K.A., Valero, J., Papsin, B.C., 2007a. Auditory brainstem activity in children with 9–30 months of bilateral cochlear implant use. *Hear Res.* 233 (1–2), 97–107.
- Gordon, K.A., Valero, J., Papsin, B.C., 2007b. Binaural processing in children using bilateral cochlear implants. *Neuroreport* 18 (6), 613–617.
- Gordon, K.A., Papsin, B.C., Harrison, R.V., 2007c. Auditory brainstem activity and development evoked by apical versus basal cochlear implant electrode stimulation in children. *Clin. Neurophysiol.* 118 (8), 1671–1684.
- Gordon, K.A., Wong, D.D., Papsin, B.C., 2013. Bilateral input protects the cortex from unilaterally-driven reorganization in children who are deaf. *Brain* 136 (Pt 5), 1609–1625.
- Gordon, K., Henkin, Y., Kral, A., 2015. Asymmetric hearing during development: the aural preference syndrome and treatment options. *Pediatrics* 136 (1), 141–153.
- Goupell, M.J., et al., 2013. Effect of mismatched place-of-stimulation on the salience of binaural cues in conditions that simulate bilateral cochlear-implant listening. *J. Acoust. Soc. Am.* 133 (4), 2272–2287.
- Graham, J., et al., 2009. Bilateral sequential cochlear implantation in the congenitally deaf child: evidence to support the concept of a 'critical age' after which the second ear is less likely to provide an adequate level of speech perception on its own. *Cochlear Implants Int.* 10 (3), 119–141.
- Grieco-Calub, T.M., Litovsky, R.Y., 2010. Sound localization skills in children who use bilateral cochlear implants and in children with normal acoustic hearing. *Ear Hear.* 31 (5), 645–656.
- Grothe, B., Pecka, M., McAlpine, D., 2010. Mechanisms of sound localization in mammals. *Physiol. Rev.* 90 (3), 983–1012.
- Hartmann, W.M., Raked, B., 1989. On the minimum audible angle—a decision theory approach. *J. Acoust. Soc. Am.* 85 (5), 2031–2041.
- He, S., Brown, C.J., Abbas, P.J., 2010. Effects of stimulation level and electrode pairing on the binaural interaction component of the electrically evoked auditory brain stem response. *Ear Hear.* 31 (4), 457–470.
- He, S., Brown, C.J., Abbas, P.J., 2012. Preliminary results of the relationship between the binaural interaction component of the electrically evoked auditory brainstem response and interaural pitch comparisons in bilateral cochlear implant recipients. *Ear Hear.* 33 (1), 57–68.
- Henkin, Y., Swead, R.T., Roth, D.A., Kishon-Rabin, L., Shapira, Y., Migirov, L., Hildesheimer, M., Kaplan-Neeman, R., 2014. Evidence for a right cochlear implant advantage in simultaneous bilateral cochlear implantation. *Laryngoscope* 124, 1937–1941.
- Illg, A., et al., 2013. Speech comprehension in children and adolescents after sequential bilateral cochlear implantation with long interimplant interval. *Otol. Neurotol.* 34 (4), 682–689.
- Jiwani, S., Papsin, B.C., Gordon, K.A., 2013. Central auditory development after long-term cochlear implant use. *Clin. Neurophysiol.* 124 (9), 1868–1880.
- Jiwani, S., Papsin, B.C., Gordon, K.A., 2016. Early unilateral cochlear implantation promotes mature cortical asymmetries in adolescents who are deaf. *Hum. Brain Mapp* 37, 135–152.
- Jones, G.L., Litovsky, R.Y., 2011. A cocktail party model of spatial release from masking by both noise and speech interferers. *J. Acoust. Soc. Am.* 130 (3), 1463–1474.
- Kan, A., et al., 2013. Effect of mismatched place-of-stimulation on binaural fusion and lateralization in bilateral cochlear-implant users. *J. Acoust. Soc. Am.* 134 (4), 2923–2936.
- Kan, A., Litovsky, R.Y., 2015. Binaural hearing with electrical stimulation. *Hear Res.* 322, 127–137.
- Kan, A., Litovsky, R.Y., Goupell, M.J., 2015. Effects of interaural pitch matching and auditory image centering on binaural sensitivity in cochlear implant users. *Ear Hear* 36, e62–e68.
- Killan, C.F., Killan, E.C., Raine, C.H., 2015. Changes in children's speech discrimination and spatial release from masking between 2 and 4 years after sequential cochlear implantation. *Cochlear Implants Int.* 16 (5), 270–276.
- Kral, A., 2013. Auditory critical periods: a review from system's perspective. *Neuroscience* 247, 117–133.
- Kral, A., et al., 2013. Single-sided deafness leads to unilateral aural preference within an early sensitive period. *Brain* 136 (Pt 1), 180–193.
- Kral, A., Sharma, A., 2012. Developmental neuroplasticity after cochlear implantation. *Trends Neurosci.* 35 (2), 111–122.
- Laback, B., Egger, K., Majdak, P., 2015. Perception and coding of interaural time differences with bilateral cochlear implants. *Hear Res.* 322, 138–150.
- Lieu, J.E., 2013. Unilateral hearing loss in children: speech-language and school performance. *B-ENT (Suppl. 21)*, 107–115.
- Lieu, J.E., et al., 2010. Unilateral hearing loss is associated with worse speech-language scores in children. *Pediatrics* 125 (6), e1348–55.
- Lieu, J.E., Tye-Murray, N., Fu, Q., 2012. Longitudinal study of children with unilateral hearing loss. *Laryngoscope* 122 (9), 2088–2095.
- Litovsky, R.Y., 1997. Developmental changes in the precedence effect: estimates of minimum audible angle. *J. Acoust. Soc. Am.* 102 (3), 1739–1745.
- Litovsky, R.Y., 2005. Speech intelligibility and spatial release from masking in young children. *J. Acoust. Soc. Am.* 117 (5), 3091–3099.
- Litovsky, R.Y., 2011. Review of recent work on spatial hearing skills in children with bilateral cochlear implants. *Cochlear Implants Int.* 12 (Suppl. 1), S30–S34.
- Litovsky, R., 2015. Development of the auditory system. *Handb. Clin. Neurol.* 129, 55–72.
- Litovsky, R.Y., et al., 2010. Effect of age at onset of deafness on binaural sensitivity in electric hearing in humans. *J. Acoust. Soc. Am.* 127 (1), 400–414.
- Litovsky, R.Y., et al., 2012. Studies on bilateral cochlear implants at the University of Wisconsin's binaural hearing and speech Laboratory. *J. Am. Acad. Audiol.* 23 (6), 476–494.
- Litovsky, R.Y., Johnstone, P.M., Godar, S., Agrawal, S., Parkinson, A., Peters, R., Lake, J., 2006. Bilateral cochlear implants in children: localization acuity measured with minimum audible angle. *Ear Hear.* 27 (1), 43–59.
- Lopez-Torrijos, M., Mengual-Andres, S., Estelles-Ferrer, R., 2015. Clinical and log-opaque results of simultaneous and sequential bilateral implants in children with severe and/or profound bilateral sensorineural hearing loss: a literature review. *Int. J. Pediatr. Otorhinolaryngol.* 79 (6), 786–792.
- Macpherson, E.A., Middlebrooks, J.C., 2002. Listener weighting of cues for lateral angle: the duplex theory of sound localization revisited. *J. Acoust. Soc. Am.* 111 (5 Pt 1), 2219–2236.
- Middlebrooks, J.C., Green, D.M., 1991. Sound localization by human listeners. *Annu. Rev. Psychol.* 42, 135–159.
- Misurelli, S.M., Litovsky, R.Y., 2012. Spatial release from masking in children with normal hearing and with bilateral cochlear implants: effect of interferer asymmetry. *J. Acoust. Soc. Am.* 132 (1), 380–391.
- Misurelli, S.M., Litovsky, R.Y., 2015. Spatial release from masking in children with bilateral cochlear implants and with normal hearing: effect of target-interferer similarity. *J. Acoust. Soc. Am.* 138 (1), 319.
- Mok, M., et al., 2007. Spatial unmasking and binaural advantage for children with normal hearing, a cochlear implant and a hearing aid, and bilateral implants. *Audiol. Neurotol.* 12 (5), 295–306.
- Mok, M., et al., 2010. Speech perception benefit for children with a cochlear implant and a hearing aid in opposite ears and children with bilateral cochlear implants. *Audiol. Neurotol.* 15 (1), 44–56.
- Nittrouer, S., Chapman, C., 2009. The effects of bilateral electric and bimodal electric–acoustic stimulation on language development. *Trends Amplif.* 13 (3), 190–205.
- Peters, B.R., et al., 2007. Importance of age and postimplantation experience on speech perception measures in children with sequential bilateral cochlear implants. *Otol. Neurotol.* 28 (5), 649–657.
- Polonenko, M.J., Papsin, B.C., Gordon, K.A., 2015. The effects of asymmetric hearing on bilateral brainstem function: findings in children with bimodal (electric and acoustic) hearing. *Audiol. Neurotol.* 20 (Suppl. 1), 13–20.
- Ponton, C.W., Eggermont, J.J., 2001. Of kittens and kids: altered cortical maturation following profound deafness and cochlear implant use. *Audiol. Neurotol.* 6 (6), 363–380.
- Probst, E.J., Greinwald, J.H., Schmithorst, V., 2010. Neuroanatomic differences in children with unilateral sensorineural hearing loss detected using functional magnetic resonance imaging. *Arch. Otolaryngol. Head. Neck Surg.* 136 (1), 22–26.
- Ramsden, J.D., Papsin, B.C., Leung, R., James, A., Gordon, K.A., 2009. Bilateral simultaneous cochlear implantation in children: our first 50 cases. *Laryngoscope* 119, 2444–2448.
- Salloum, C.A., et al., 2010. Lateralization of interimplant timing and level differences in children who use bilateral cochlear implants. *Ear Hear.* 31 (4), 441–456.
- Scherf, F., et al., 2009a. Subjective benefits of sequential bilateral cochlear implantation in young children after 18 months of implant use. *ORL J. Otorhinolaryngol. Relat. Spec.* 71 (2), 112–121.
- Scherf, F., et al., 2009b. Three-year postimplantation auditory outcomes in children with sequential bilateral cochlear implantation. *Ann. Otol. Rhinol. Laryngol.* 118 (5), 336–344.
- Schmithorst, V.J., et al., 2005. Cortical reorganization in children with unilateral sensorineural hearing loss. *Neuroreport* 16 (5), 463–467.
- Schmithorst, V.J., Plante, E., Holland, S., 2014. Unilateral deafness in children affects development of multi-modal modulation and default mode networks. *Front. Hum. Neurosci.* 8, 164.
- Smith, Z.M., Delgutte, B., 2007. Using evoked potentials to match interaural electrode pairs with bilateral cochlear implants. *J. Assoc. Res. Otolaryngol.* 8 (1), 134–151.
- Sparreboom, M., et al., 2010. Electrically evoked auditory brainstem responses in children with sequential bilateral cochlear implants. *Otol. Neurotol.* 31 (7), 1055–1061.

- Sparreboom, M., Snik, A.F., Mylanus, E.A., 2011. Sequential bilateral cochlear implantation in children: development of the primary auditory abilities of bilateral stimulation. *Audiol. Neurotol.* 16, 203–213.
- Steel, M.M., Papsin, B.C., Gordon, K.A., 2015. Binaural fusion and listening effort in children who use bilateral cochlear implants: a psychoacoustic and pupillometric study. *PLoS One* 10 (2), e0117611.
- Strom-Roum, H., et al., 2012. Sound localising ability in children with bilateral sequential cochlear implants. *Int. J. Pediatr. Otorhinolaryngol.* 76 (9), 1245–1248.
- Summerfield, A.Q., et al., 2002. A cost-utility scenario analysis of bilateral cochlear implantation. *Arch. Otolaryngol. Head. Neck Surg.* 128 (11), 1255–1262.
- Thai-Van, H., et al., 2007. The pattern of auditory brainstem response wave V maturation in cochlear-implanted children. *Clin. Neurophysiol.* 118 (3), 676–689.
- Tibbetts, K., et al., 2011. Interregional brain interactions in children with unilateral hearing loss. *Otolaryngol. Head Neck Surg.* 144 (4), 602–611.
- Todd, A.E., Goupell, M.J., Litovsky, R.Y., 2015. Binaural release from masking with single- and multi-electrode stimulation in children with cochlear implants. *J. Acoust. Soc. Am.* (submitted for publication).
- Van Deun, L., et al., 2011. Binaural unmasking of multi-channel stimuli in bilateral cochlear implant users. *J. Assoc. Res. Otolaryngol.* 12 (5), 659–670.
- Van Deun, L., van Wieringen, A., Wouters, J., 2010. Spatial speech perception benefits in young children with normal hearing and cochlear implants. *Ear Hear.* 31 (5), 702–713.
- Vasama, J.P., Makela, J.P., 1997. Auditory cortical responses in humans with profound unilateral sensorineural hearing loss from early childhood. *Hear Res.* 104 (1–2), 183–190.
- Vincent, C., et al., 2015. Identification and evaluation of cochlear implant candidates with asymmetrical hearing loss. *Audiol. Neurotol.* 20 (Suppl. 1), 87–89.
- Wolfe, J., et al., 2007. 1-year postactivation results for sequentially implanted bilateral cochlear implant users. *Otol. Neurotol.* 28 (5), 589–596.
- Yin, T.C.T., 2002. Neural mechanisms of encoding binaural localization cues in the auditory brainstem. In: Oertel, D., Fay, R.R., Popper, A.N. (Eds.), *Springer Handbook of Auditory Research – Integrative Functions in the Mammalian Auditory Pathway*. Springer-Verlag, New York, pp. 99–159.
- Zheng, Y., Godar, S.P., Litovsky, R.Y., 2015. Development of sound localization strategies in children with bilateral cochlear implants. *PLoS One* 10 (8), e0135790.